

SPECIFICATIONS:

The Amico Regal Series Flatwall shall be manufactured by Amico Corporation, in accordance with these shop drawings and documents. The complete flat wall shall be UL and cUL listed.

1. Basic Construction

Enclosure shall be constructed of extruded, anodized aluminum alloy sections to provide a modular surface mounted unit. The unit shall be factory assembled to include mechanical and electrical components as shown on the plans and specifications and wired in accordance with Amico Corporation’s shop drawings as herein indicated. Anodized aluminum facials shall be removable for access to individual components mounted within the console sections of the wall unit. Primary electrical conduit runs to the unit are provided by the electrical contractor and shall be accommodated by an upper terminal enclosure furnished as an integral part of the wall. All secondary wiring within the wall unit enclosure shall be enclosed in electrical conduits. Each side of the enclosure provided by Amico Corporation and terminated in the appropriate sections in the upper terminal enclosures. All wiring shall be marked, color-coded to facilitate easy circuit identification.

Removable panels shall be finished with high-pressure plastic laminate on the outside face and a backing sheet on the reverse. The color of the plastic laminate shall be selected by the hospital or architect from standard color selections from Formica or Wilson Art.

2. Components

Medical gas station outlets. Medical gas outlets shall be as indicated in plans and specifications. Amico Corporation will factory install all outlets, manifolded and tested in accordance with the outlet manufacturer’s requirements. Each outlet, piping and manifolding shall be pressure tested for 24 hours. The mechanical contractor will bring the primary supply lines of the medical gases to the singular terminal connections at locations shown on the drawings.



Power outlets. Receptacles shall be the type and quantity as shown on the drawings and specifications. 120 Volt, Hospital Grade shall be provided.

Chart lights / Night Light. Shall be switched, (where required), 4 watts, 120 VAC to provide light for reading records in a darkened room. The front of the light shall be louvered directed downward and located in front or side of the wall. It shall be complete with bulb.

Vacuum Bottle Slide. Shall be a 4 inch anodized aluminum device with bottom stop. It shall be installed on the wall as specified on drawings.

Internal ground bus. A screw type ground bus connection will be provided as required. The ground bus assembly shall also provide a positive means for grounding the wall.

Switches. Shall be as shown on drawings, 120 VAC 15 or 20 Amp “Specification” Grade.

Grounding jacks. Externally accessible grounding jacks shall be provided in the flat wall unit, as shown in the shop drawings. The solid brass receptacle shall be enclosed in a non-conductive housing, shall be spring-loaded and shall be manufactured to the requirements of NFPA 99.

PROJECT:

Lighting fixture. When specified, the over-bed lighting fixture shall be provided complete with plastic laminate trim color to match the flat wall unit. Amico Corporation does not provide lamps.

Clock timer. When specified, an Amico Corporation Clock/Timer unit shall be provided and factory installed in the flat wall. The microprocessor based unit displays both time-of-day and elapsed time. The clock shall have an internal backup circuit that keeps the clock and timer functions running for at least 24 hours. The Clock/Timer unit shall have a provision for Code Blue input. The installing contractor shall perform wiring and electrical actuation between the Code Blue and the Clock/Timer unit.

3. Provisions

Nurse call equipment. All nurse call equipment shall be furnished and installed by others as indicated in the plans and specifications. Amico Corporation shall provide a pull cord for use by the nurse call installer.

Code Blue. Code Blue equipment shall be furnished and installed by others as indicated in the plans and specifications. Amico Corporation shall provide a pull cord for use by contractor.

Monitoring equipment. Monitor equipment shall be provided and installed by others. Amico Corporation shall provide a matching blank plate for the monitor receptacle to the monitor supplier when requested. The wall unit will include a factory-installed raceway and labeled pull cord to accommodate the monitor wiring. Where specified, a Tilt Swivel Monitor Bracket and slide shall be provided to support the monitor.

Data/Computer interconnects. A data/computer system furnished by others shall be installed at locations as shown on the plans and specifications. Amico Corporation shall provide a matching blank plate for the computer-jack when requested. The wall unit will include a factory-installed raceway and pull cord to accommodate the computer system wiring.

Telephone Provisions. Shall include a 1/2" conduit with "Pull-cord" to junction box. Telephone Jack shall be installed and wired by contractor or local telephone company.

4. Accessories

Accessories shall be provided in the quantities and types as shown in the submittal. Accessories shall be installed by others in the locations as determined by the end user.

5. Installation

Mounting of the product shall be the responsibility of the installing contractor. This includes receiving, storage, erection, overhead bracing, clean up, touch up, carton disposal etc. This contractor shall supply all necessary installation material including such items as tools, fasteners, caulking and electric lamps not supplied by Amico Corporation.

The electrical contractor shall be responsible for all electrical hook-up at service connection locations plus interconnect wiring on multi-section units. After the installation is complete the electrical contractor shall test equipment function plus electrical receptacles and grounding in accordance with local electrical code requirements.

The medical gas contractor shall be responsible for piping and hook-up of all medical gas services. In addition, hook-up of piping between sections on multi-section units shall be the responsibility of this contractor. The medical gas contractor shall be responsible for purging, pressure testing, gas identification and system certification in accordance with NFPA 99 or CSA-Z305.1.

Accessory items shall be installed in accordance with the manufacturer's instructions and shall be the responsibility of this section. The installation shall take place under the direction of the hospital. Please see our accessories catalog for additional information.

Note: All Specifications are nominal and subject to change without notice.

D I S T R I B U T E D B Y :